

Quality Control Certificate

Instrument

MODEL/CAT. NO. : ACURA 825.0200
VOLUME RANGE : 20-200 μ l
SERIAL NO. 13122016 NEW

Test data

CONTROL DATE : 19-Jan-2004
OPERATOR ID MO
BALANCE checked against Swiss master standard
AIR TEMP : Between 20-25°C [\pm 0.5°C]
AIR PRESSURE : 1025 hPa
WATER TEMP : 22 °C
LIQUID DENSITY 0.997768 g/cm³
METHOD : Based on DIN 12650; ISO 8655

Test results

VOLUME	MEAN	INACCURACY	IMPRECISION
20 μ l	19.99 μ l	E -0.01 μ l E -0.05 %	SD 0.10 μ l CV 0.50 %
Status :		passed	passed
200 μ l	199.91 μ l	E -0.09 μ l E -0.05 %	SD 0.35 μ l CV 0.18 %
Status :		passed	passed

E=Mean error SD=Standard deviation CV=Coefficient of variation

Warranty Conditions

This product is guaranteed for a three year period against any manufacturing or material defect. Cases due to the non-respect of operating and safety instructions as well as colour alterations are excluded from the warranty. Claim return authorisation from distributor from whom it was purchased prior to returning it for repair. Clean and decontaminate instrument before shipping.



SOCOREX



SOCOREX ISBA SA - CHAMP-COLOMB 7 - P.O. BOX
1024 ECUBLENS/LAUSANNE - SWITZERLAND
FAX +41(0)21 651 6001 - e-mail: socorex@socorex.com - www.socorex.com